



SuperSystems

incorporated



Basic Nitriding Sampling System Hydrogen Analyzer with Calculated % DA, % NH₃, and K_N Values

Operations Manual

Please read, understand, and follow these instructions before operating this equipment. Super Systems, Inc. is not responsible for damages incurred due to a failure to comply with these instructions. If at any time there are questions regarding the proper use of this analyzer, please contact us at 513-772-0060 for assistance.

Super Systems Inc.

USA Office

Corporate Headquarters:

7205 Edington Drive

Shipping Address:

7245 Edington Drive

Cincinnati, OH 45249

Phone: (513) 772-0060

<http://www.supersystems.com>

Super Systems Europe

Units 3 & 4, 17 Reddicap Trading Estate,

Sutton Coldfield, West Midlands

B75 7BU

UNITED KINGDOM

Phone: +44 (0) 121 329 2627

<http://www.supersystemseurope.com>

Super Systems México

Sistemas Superiores Integrales S de RL de CV

Calle 3 Int.: 11.

Zona Ind. Benito Juarez

Querétaro, Qro. Méx.

C.P.: 76120

Phone: +52 (442) 210 2459

<http://www.supersystems.com.mx>

Super Systems China

No. 335 XianXia Road

Room 308

Shanghai, CHINA

200336

Phone: +86 21 5206 5701/2

<http://www.supersystems.com>

Table of Contents	
Introduction	4
Specifications	4
Electrical Connections / Terminal Assignments	4
Getting Started	4
Default Settings	5
Modifying the Default Settings	5
Description of Menu Items	5
Exit Program	5
Instrument Information	5
Communication Setup	5
Instrument Configuration	5
Percent Hydrogen	5
Percent Dissociation	6
Percent Ammonia	6
Nitriding Potential	6
Output Configuration	6
Output Calibration	6
Sensor Calibration	6
Performing a Zero Calibration	7
Performing a Span Calibration	7
Parts List and Internal Components	7
Warranty	9
Revision History	10
Appendix A: Drawings (Electrical, Plumbing, and Mounting)	11

Introduction

This instrument uses the measurement of Hydrogen to display % Hydrogen (H₂), % Dissociation (DA), and % Ammonia (NH₃). When the flow rates of Nitrogen, Ammonia, and Dissociated Ammonia are manually entered, the instrument can also calculate Nitriding Potential (K_N).

Specifications

- Power Requirements: 100-240 VAC
- Current Draw: Max. 0.2 Amps
- Sensor Technology: Thermal Conductivity
- User Interface: 3.5" Color QVGA TFT LCD Touch Screen
- Measurement Range: 0-100% H₂
- Hydrogen Measurement Accuracy: +/- 1%
- Hydrogen Measurement Repeatability: +/- 1%
- Hydrogen Measurement Resolution: 0.01%
- Analog Outputs: Two Isolated 4-20mA (User Configurable)
- Analog Output Resolution: 0.005mA
- Analog Output Accuracy: +/- 0.01% of Range
- Analog Output Linearity: +/- 0.01%
- Analog Output Load Resistance: Minimum 0 Ohm, Maximum 500 Ohm
- Digital Communications: RS485 Modbus, Ethernet
- Enclosure Size (Without Filter): Approx. 13"L x 13"W x 6"D / 330mmL x 330mmW x 142mmD
- Enclosure Weight: 22.5 lbs. / 10.2 kg.
- Enclosure Ventilation: Continuous Purge Fan, Dual Vents
- Recommended Flow Rate: 1.5 to 2.0 SCFH / 0.71 to 0.94 lpm
- Process Gas Fittings: Stainless Steel Compression for ¼" OD Tubing
- Calibration Gas Fitting: 1/8" Barb (can be removed for 1/8" Female NPT Port)
- Operating Environment: 10-90 %RH (Non-Condensing)
- Operating Temperature: 32 to 122°F / 0 to 50°C
- Sample Gas Temperature: 32 to 158°F / 0 to 70°C

Electrical Connections / Terminal Assignments

Wire Number	Function
1000	AC Line Power (100-240VAC)
1002	AC Neutral Power
Ground	AC Ground
1121	Analog Output Common (+)
1111	Analog Output #1 (-)
1121	Analog Output Common (+)
1131	Analog Output #2 (-)
1081	RS485 (-)
1091	RS485 (+)

Getting Started

Please refer to the enclosed drawing for instructions regarding the proper electrical and mechanical installation of this instrument. The flow meter on the door of the enclosure should be adjusted to maintain 1.5 SCFH of process gas flow.

Default Settings

When the instrument is turned on, it will display % H₂ on the screen. To display additional values (% DA, % NH₃, or K_N), see the “Instrument Setup” screen. The first 4-20mA output will be set up to retransmit the %H₂ value scaled for 0 to 100%, and the second output will be set up for %DA also scaled for 0-100%. Any changes to these default parameters will be stored so they will not need to be re-entered when the power is shut off to the instrument

Modifying the Default Settings

To access the operational and setup parameters, press the “menu” button at the lower left section of the screen. This will allow you to select only two options, Exit Program and Instrument Information. To prevent unwanted modification to the operation of the instrument, these are the only options available unless the user logs in. To access the other menu options, press the “Login” button and enter “2”. This will give the operator access to the setup and operational parameters.

Each individual menu option is described in detail below:

Description of Menu Items

Exit Program

The touch screen display is constantly writing data to its flash card for storage, and it is important that the instrument not be shut down during this process. Before removing power from the enclosure, select “Exit Program”. It will ask if you are sure that you want to shut down the interface. By answering yes, it closes the operating program in an orderly manner. Once the standard Microsoft Windows screen appears it is safe to remove power from the instrument. If this procedure is not followed, there is a chance that there could be an error writing to the flash card that could cause problems with the operation of the instrument.

Instrument Information

This screen provides information on any applicable revision levels and serial numbers. It also shows if the instrument is logging data. There are no functions that can be performed on this screen; it is for informational purposes only.

Communication Setup

The communication methods shown on this screen are for display only and cannot be modified. The baud rates can be adjusted but they have been optimized for this instrument and modification is not recommended.

Instrument Configuration

This instrument is capable of displaying four different parameters. These include:

- Percent Hydrogen (H₂)
- Percent Dissociation (DA)
- Percent Ammonia (NH₃)
- Nitriding Potential (K_N)

The Percent Hydrogen is displayed on the main screen at all times. One additional parameter can be displayed along with Hydrogen by highlighting it and pressing the “OK” button. A clearer description of each of these parameters can be found in Appendix “A”, Gas Nitriding Technical paper.

Percent Hydrogen

The percent hydrogen is the amount of Hydrogen that is being detected by the thermal conductivity sensor inside the instrument. There are no additional calculations being performed to this value.

Percent Dissociation

Dissociation is derived from the amount of Hydrogen in the sample gas.

Percent Ammonia

The amount of Ammonia can also be inferred from the Hydrogen value.

Nitriding Potential

The accurate calculation of Nitriding Potential (K_N) requires the flow rates of other gases that are being introduced into the process. These flow rates are measured in SCFH (Standard Cubic Feet per Hour). K_N can only be calculated by entering the flow of Nitrogen (N_2), Ammonia (NH_3) and % Dissociated Ammonia (%DA). These values are entered at the bottom of the main screen when K_N is selected.

Output Configuration

There are two 4-20mA outputs that can be configured for any of the four parameters. For each input, the operator can select the Source (H_2 , DA, NH_3 , K_N), the zero value (the value to be represented by 4mA) and the span value (the value to be represented by 20mA).

- Output 1 can be measured from Terminal #1121(+) and 1111 (-).
- Output 2 can be measured from Terminal #1121(+) and 1131 (-).

Output Calibration

Accurate retransmission of the selected parameters can only be obtained through calibrating both of the analog outputs. This is done at the factory prior to shipment, however it is a relatively simple procedure that can be performed in the field if desired. To perform a calibration, a multimeter with a current input is required. Please keep in mind that the accuracy of the instrument used to calibrate the outputs is directly related to the accuracy of the outputs after calibration, so a meter calibrated against NIST (National Institute of Standards and Technology) standards is preferred. Before performing any calibrations, all other devices must be disconnected from the analog outputs. Multiple devices on the outputs will cause inaccurate measurement of current.

Attach the leads of the multimeter to the terminals for Output 1. The positive lead should be attached to Terminal #1121 and the negative lead to Terminal 1111. Select "Zero Output 1" and press the "Prep for Cal" button. The low limit of the output is 4mA, so 4.000 is shown as the default measured value. Ideally the meter connected to the outputs will also show 4.000. If the two values are not close enough to obtain the desired level of accuracy, a calibration should be performed.

If, for example, the meter connected to the outputs reads 4.216mA, then the value "4.216" should be entered on the screen as the Measured Value. Once the value has been entered, press the "calibrate" button. This will offset the mA output of the instrument in an amount that results in an exact output of 4.000mA. When the calibration is complete, the multimeter should be reading 4.000mA (+/- .005).

The procedure can be repeated for the Span of Output 1. When "Span Output 1" is selected and "Prep for Cal" is pressed, the instrument will output the high output limit, which is 20mA. The multimeter will display the actual output from the instrument, and if it is not within the desired tolerance it can be calibrated using the same procedure as above.

To calibrate Output 2, attach the leads of the multimeter to Terminal #1121 (+) and the negative lead to Terminal 1131 (-) and follow the same procedure that was used for the first output.

Sensor Calibration

A proper calibration of the sensor requires two gases. The first gas should be pure Nitrogen or Argon. This contains no Hydrogen, and is therefore referred to as the Zero Gas. The second gas is the Span Gas. The Span Gas should ideally contain a quantity of Hydrogen similar to the amount of Hydrogen in the process gas. The Span Gas should also include any other gases that are present in the process gas in

their respective percentages. The more similar to the process gas the calibration gas is, the more accurate the calibration will be.

Performing a Zero Calibration

On the Sensor Calibration page, select “Zero Hydrogen”. Turn the valve on the side of the enclosure to the “Calibration Gas” position, and attach the Zero gas to the “Calibration Gas Inlet” port. Begin the flow of gas at a rate of 1.5 to 2.0 SCFH as measured on the flow meter on the door of the enclosure. The gas should not be under any pressure other than the amount required to maintain the appropriate flow amount. The target Value is shown on the screen. For a Zero Calibration, this will be 0.00 (the amount of Hydrogen in the Zero Gas). The Measured H₂ Value can be seen at the bottom of the screen. When this value comes to equilibrium, it will not be showing any upward or downward trends, only the slight oscillation of the readings. This usually occurs in approximately 30 seconds. When the sensor is at equilibrium, press the green “Calibrate” button to perform the zero calibration. After the Zero Calibration is complete, turn off the flow of gas and disconnect it from the enclosure.

Performing a Span Calibration

To perform a span calibration, select “Span Hydrogen”, attach the Span Gas to the Calibration Gas Inlet port, and begin the flow of gas at 1.5 to 2.0 SCFH. The Target Value should be set to the exact amount of Hydrogen that is in the Span Gas cylinder. Then the same procedure should be followed as the Zero calibration, with the “Calibrate” button being pressed after the readings reach equilibrium. After the Span Gas calibration is complete, turn off the flow of gas, disconnect the cylinder from the enclosure, and restore the valve on the side of the instrument to the “Sample Gas” position. This will re-connect the sensor to the process gas stream.


Parts List and Internal Components

The following items can be purchased as needed for the Hydrogen Nitrider Analyzer.

Part Number	Description
31604	2 GB SD card
37026	Fan (24VDC, 60mm)
37029	Fan filter assembly
37150	3-way ball valve
37050	High capacity bowl filter
37051	High capacity filter element
34444	Tubing
33097	Circuit breaker
31135	Power supply
A20624	Hydrogen sensor
31296	3.5” touch screen
Full Unit	
13537	Hydrogen Nitrider Analyzer

Basic Nitriding Sampling System – Hydrogen Analyzer

The following diagram illustrates the location of important internal components of the Hydrogen Nitrider Analyzer, along with relevant part numbers.



*SD card is contained within touch screen.
SSi part number for 2 GB SD card is 31604.

**Fan filter assembly (PN 37029) is contained within fan housing.

Warranty

Limited Warranty for Super Systems Products:

The Limited Warranty applies to new Super Systems Inc. (SSI) products purchased direct from SSI or from an authorized SSI dealer by the original purchaser for normal use. SSI warrants that a covered product is free from defects in materials and workmanship, with the exceptions stated below.

The limited warranty does not cover damage resulting from commercial use, misuse, accident, modification or alteration to hardware or software, tampering, unsuitable physical or operating environment beyond product specifications, improper maintenance, or failure caused by a product for which SSI is not responsible. There is no warranty of uninterrupted or error-free operation. There is no warranty for loss of data—you must regularly back up the data stored on your product to a separate storage product. There is no warranty for product with removed or altered identification labels. SSI DOES NOT PROVIDE ANY OTHER WARRANTIES OF ANY KIND, INCLUDING, BUT NOT LIMITED TO, THE IMPLIED WARRANTIES OR CONDITIONS OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. SOME JURISDICTIONS DO NOT ALLOW THE LIMITATION OF IMPLIED WARRANTIES, SO THIS LIMITATION MAY NOT APPLY TO YOU. SSI is not responsible for returning to you product which is not covered by this limited warranty.



If you are having trouble with a product, before seeking limited warranty service, first follow the troubleshooting procedures that SSI or your authorized SSI dealer provides.

SSI will replace the PRODUCT with a functionally equivalent replacement product, transportation prepaid after PRODUCT has been returned to SSI for testing and evaluation. SSI may replace your product with a product that was previously used, repaired and tested to meet SSI specifications. You receive title to the replaced product at delivery to carrier at SSI shipping point. You are responsible for importation of the replaced product, if applicable. SSI will not return the original product to you; therefore, you are responsible for moving data to another media before returning to SSI, if applicable. Data Recovery is not covered under this warranty and is not part of the warranty returns process. SSI warrants that the replaced products are covered for the remainder of the original product warranty or 90 days, whichever is greater.

Revision History

Rev.	Description	Date	MCO#
NEW	First release		N/A
A	Revised terminal numbers; added product diagrams; added part numbers; updated format of document.	11/7/2014	2155
B	Revised specifications for hydrogen measurement.	9/23/2015	2168

Appendix A: Drawings (Electrical, Plumbing, and Mounting)



Basic Nitriding Sampling System – Hydrogen Analyzer

